

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Examiner Lyle A. Alexander
ART UNIT 1743

In re application of
E. Alan Bates et al.
Application No. 08/935,629
Filed 09/23/97
For ASSAYING DEVICE

DECLARATION UNDER 37 CFR 1.132

Commissioner for Patents
Washington, DC 20231

I, Gary Hoffman, declare as follows:

1. I am the third joint inventor for this application.

2. While the claims of this application have been rejected based on the patent of Senior, the disclosure of the application starts from a technology actually quite different from that of Senior. Characteristic of this difference is Senior's provision of its bibulous member 16 protruding out of the housing. When the bibulous member is placed in a urine stream, it is quickly soaked. Senior has correctly chosen to display this version in its drawings, because versions such as suggested in paragraph a. in col. 5 of Senior would tend to shield the bibulous member from a urine stream.

3. This application references in its BACKGROUND a quite different technology, that of drug testing using an immunoassay method called antigen-antibody competitive binding. Characteristic of this technology is the measured application of only a few drops of urine. This dropwise application of the urine is disclosed in the specification in the first paragraph of the DETAILED DESCRIPTION OF THE INVENTION.

4. Attached hereto are Exhibits 1 to 5 demonstrating five different instances of the type

of technology forming the starting point for this application. These are as follows:

Exhibit 1 - Pages 1-4 of a document entitled "AccuSign DOA 4, THC/OPI/COC/AMP" bearing copyright notice dated 1996;

Exhibit 2 - Pages 1-4 of a document entitled "AccuSign BAR", also bearing copyright notice dated 1996;

Exhibit 3 - Front and back of a leaflet headed "Drug Test Resources International", likewise concerning AccuSign DOA 4 and bearing a 1996 copyright notice;

Exhibit 4 - One-sided, undated leaflet headed "Visaline II"; and

Exhibit 5 - Copies of the packages of several kits labeled HOME DRUG TEST and their instruction leaflets.

5. Exhibit 1 is noteworthy for the correspondence of the terminology in its section Principle on its page 1 with the terminology in the BACKGROUND section of this application.

6. All of the exhibits direct that 3 drops of urine be applied, this being a standard for this technology. All of the exhibits provide either a dropper or a pipette for transfer of the urine sample into the sample well. In each of the exhibits, the sample receiving area is clearly a well, and Exhibit 1, for instance, calls it a "well" in the section Test Protocol on its page 2.

7. The chart at the top of page 2 of Exhibit 1, for instance, explains that appearance of a line for a particular drug in this starting technology is a negative indication, i.e. the drug is not present in amounts above the cutoff level.

8. A characteristic of this technology is that the sample collection location must not be flooded with urine, as in Senior, because this leads to false positives by washing out the lines that would otherwise indicate negative readings.

FROM : HOFFMAN

FAX NO. : 412-621-2420

Aug. 21 2000 02:48PM P1

9. It is to be noted that none of these exhibits mentions photocopying the results. The statements regarding photocopying in the BACKGROUND section of the present application represent perceptions of the present inventor group, rather than the state of the art at the time this invention was made.

10. While the provision of a cap to cover Senior's wet, protruding member 16 is immediately understandable, it was the present group of inventors which perceived the advantages of such for the different technology from which the present invention arose.

11. A difference between the present invention and Senior concerns the problem with putting a cap on Senior's test. Because the bibulous material can be so flexible (especially when wet) there often is difficulty slipping the cap on without bending the material. It's like threading a needle by holding the thread still and moving the needle.

12. All statements made herein of my own knowledge are true and all statements made on information and belief are believed to be true; such statements were made with the knowledge that willful false statements and the like so made are punishable by fine or imprisonment, or both, under Section 1001 of Title 18 of the United States Code and may jeopardize the validity of the application or any patent issued thereon.

Signature: Gary A. [Signature]Date: 8-21-00

AccuSign™ DOA 4

THC/OPI/COC/AMP

One-Step Panel Assay for Drugs of Abuse

For In Vitro Use Only

Simple One-Step Immunoassay for the
Qualitative Detection of THC metabolites,
Opiates, Cocaine metabolite, Amphetamines,
and/or their Metabolites in Urine

PBM

Catalog No.	DOA-240	35 Test Kit
	DOA-240-10	10 Test Kit

Intended Use

The AccuSign™ DOA 4 THC/OPI/COC/AMP Panel Assay is a simple, one-step immunochemical test for the rapid, qualitative detection of THC metabolites, opiates, cocaine metabolite, and amphetamines in urine. The test detects the major metabolites of these drugs at the following cutoff concentrations.

THC	11-nor- Δ^9 -THC-9-carboxylic acid	50 ng/mL
OPI	Morphine	300 ng/mL
COC	Benzoyllecgonine	300 ng/mL
AMP	Amphetamine	1000 ng/mL

The AccuSign™ DOA 4 THC/OPI/COC/AMP test provides only a preliminary analytical result. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography, mass spectrometry (GC/MS) is the preferred confirmatory method. Other chemical confirmatory methods are available. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.

Summary and Explanation

Drug abuse has become one of the most destructive social problems in recent years, affecting nearly every corner of the world. To effectively combat this increasingly disturbing problem, there is a strong need for a simple, rapid, inexpensive, disposable, visual, and non-instrument requiring drug screening test kit. According to the National Institute on Drug Abuse (NIDA), THC (Marijuana), Opiates, Cocaine, and Amphetamines are among the most widely abused drugs. The one-step AccuSign™ DOA 4 Panel Assay is a test for screening these four major drugs of abuse in urine, simultaneously with one sample application. The test takes less than 10 minutes to perform.

THC (Δ^9 -tetrahydrocannabinol) is the primary active ingredient in cannabinoids (marijuana). When ingested or smoked, it produces euphoric effects. Users experience impairment of short term memory and THC use slows learning. Also, it may cause transient episodes of confusion, anxiety, or frank toxic delirium. Long term, relatively heavy use may be associated with behavioral disorders. The peak

effect of smoking THC occurs in 20–30 minutes and the duration is 90–120 minutes after one cigarette. Elevated levels of urinary metabolites are found within hours of exposure and remain detectable for 3–10 days after smoking. The main metabolite excreted in the urine is 11-nor- Δ^9 -tetrahydrocannabinol-9-carboxylic acid.

Opioid analgesics comprise a large group of substances which control pain by depressing the central nervous system. Morphine is the prototype compound of this group.¹ Morphine is excreted unmetabolized, and is also the major metabolic product of codeine and heroin. Morphine is detectable in the urine for several days after an opiate dose.

Cocaine, derived from the leaves of coca plant, is a potent central nervous system (CNS) stimulant and a local anesthetic. Cocaine induces euphoria, confidence and a sense of increased energy in the user; these psychological effects are accompanied by increased heart rate, dilation of the pupils, fever, tremors and sweating. Cocaine is used by smoking, intravenous, intranasal or oral administration, and excreted in the urine primarily as benzoylecgonine in a short time. Benzoylecgonine has a longer biological half-life (5–8 hours) than cocaine (0.5–1.5 hours) and can generally be detected for 24–60 hours after cocaine use or exposure.¹⁴

Amphetamine is a potent sympathomimetic agent with therapeutic applications. The drug can be taken orally, injected, or inhaled. Acute higher doses lead to enhanced stimulation of the central nervous system and induce euphoria, alertness, reduced appetite, and a sense of increased energy and power.¹⁵ Cardiovascular responses to amphetamine include increased blood pressure and cardiac arrhythmias. More acute responses include anxiety, paranoia, hallucinations, psychotic behavior, and eventually, depression and exhaustion. The effects of amphetamine generally last 2–4 hours, and the drug has a half-life of 9–24 hours in the body. Amphetamine is excreted in the urine in unchanged form and also as hydroxylated and deaminated derivatives.¹⁶

Principle

The AccuSign™ DOA 4 test employs one-step, solid-phase immunoassay technology to discretely detect the presence of any of the above four drugs, or their immunoreactive metabolites, in urine. The assay uses highly specific monoclonal/polyclonal antibodies raised against the target drugs. The test card contains a membrane strip, on which each of the four drugs conjugated to BSA is immobilized at specific locations. The assay is based on the principle of the highly specific immunochemical reactions between antigens and antibodies which are used for the analysis of specific substances in biological fluids. The drug detection relies on the competition for binding to the antibodies between drug conjugates and drugs which may be present in the urine sample.

In the test procedure, a sample of urine is placed in the sample well of the device, and the sample is allowed to migrate upward. If any of the four drugs is present in the urine sample, it forms a complex with the antibody-dye conjugate specific for that drug, and the complex migrates toward the opposite end of the card, passing the specific locations on the membrane where each of the four drug conjugates is immobilized. The drug in the sample competes with the drug conjugate, which is immobilized on the membrane, for the limited antibodies present in the form of antibody-dye conjugate. When a sufficient amount of drug is present, the drug will saturate the antibodies, and the antibody-dye conjugate cannot bind to the drug conjugate on the membrane. Therefore, a drug-positive urine sample will not generate a line at the specific drug position in the result window, indicating a positive result from positive drug competition. Conversely, if a particular drug is absent in the urine specimen, the antibody on the antibody-dye conjugate will bind the membrane-bound drug. In this case, a drug-negative urine sample will generate a line at the specific drug position in the result window, indicating a negative result from an absence of competition with free drug.

EXHIBIT 1

AccuSign™ DOA 4

Add 3 drops (150 µL)

Read in 5–10 minutes

CONTROL (VALIDATION) LINE (C)
The Control/Validation line indicates:

1. If the proper amount of sample was used;
2. If the sample wicked;
3. If the procedure was followed properly.

If no control line appears, the test is NOT VALID. Repeat the test using a new device, and follow the procedure carefully.

OR

THC (–)
OPI (–)
COC (–)
AMP (–)

OR

THC (+)
OPI (+)
COC (+)
AMP (–)

OR

THC (–)
OPI (+)
COC (–)
AMP (–)

OR

THC (–)
OPI (–)
COC (–)
AMP (–)

INVALID

THC (–)
OPI (–)
COC (–)
AMP (–)

Negative (–) = Control line and Specific Drug line
Positive (+) = Control line only; No Specific Drug line

In addition, the test card has a procedural control built into the system, in the upper control line area. The control line is immobilized with polyclonal anti-mouse antibody; therefore, it will capture monoclonal antibody-dye conjugates that pass the region, showing a colored line in the control (validation) zone. The line works as a procedural control, confirming that proper sample volume was used and the reagent system worked. If insufficient sample volume is used, there may not be a control line, indicating the test is invalid.

Materials Provided

The AccuSign™ DOA 4 test kit contains all the reagents necessary to perform the assay.

- AccuSign™ DOA 4 device. The test device contains a membrane coated with drug conjugates in a protein matrix and a pad containing mouse monoclonal anti-THC antibody-dye conjugate, mouse monoclonal anti-opiate antibody-dye conjugate, mouse monoclonal anti-benzoylcegonine antibody-dye conjugate, and polyclonal sheep anti-amphetamine antibody-dye conjugate in a protein matrix.
- Disposable sample dispenser.
- Instructions for use.

Precautions

- For *in vitro* diagnostic use only.
- Avoid cross contamination of urine samples by using a new urine specimen container and dropper for each urine sample.
- This test kit does not contain any HIV or hepatitis infectious components. However, urine specimens are potentially infectious. Proper handling and disposal methods should be followed, according to good laboratory practices.
- The AccuSign™ device should remain in its original sealed pouch until ready for use.
- Do not use the test kit after the expiration date.

Storage and Stability

The AccuSign™ DOA 4 test kit should be stored at 2–30°C (35–86°F) in the original sealed pouch. The expiration dating was established under these storage conditions.

Specimen Collection and Preparation

Approximately 150 µL of urine sample is required for each test. Fresh urine specimens do not require any special handling or pretreatment. Specimens should be collected in a clean glass or plastic container. If

testing will not be performed immediately, specimens should be refrigerated (2–8°C) or frozen. Specimens should be brought to room temperature before testing.

Specimens containing a large amount of particulate matter may give inconsistent test results. These specimens should be clarified by centrifuging or allowing to settle before testing.

Test Procedure

The test procedure consists of adding the urine sample to the Sample well of the device and watching for the appearance of colored lines in the result window.

Test Protocol

1. For each test, open one AccuSign™ DOA 4 pouch and label the AccuSign™ device with the patient ID.
2. Holding the dropper vertically, dispense 3 full drops (150 µL) of the urine sample into the Sample well.
3. Read the result after 5–10 minutes.

Interpretation of Results

Negative: The appearance of a reddish-purple Control line (C) and a line for a specific drug indicates a negative test result; i.e., no drug above the cutoff level has been detected. The color intensities of the Control line and specific drug line may not be equal. A negative test result does not indicate the absence of drug in the sample, it indicates only that the sample does not contain drug above the cutoff level in qualitative terms.

Positive: The appearance of only a reddish-purple Control line and no distinct line next to a specific drug name indicates the test result is positive for that drug (i.e., the specimen contains the drug at a concentration above the cutoff level). A positive test result does not provide any indication of the level of intoxication or urinary concentration of the drug in the sample, it indicates only that the sample contains drug above the cutoff level in qualitative terms.

Invalid: A distinct Control line (C) should always appear. The test is invalid if no Control line forms at the C position. Such tests should be repeated with a new AccuSign™ DOA 4 test device.

Examples of possible results are shown in the diagram above.

- THC (-), Opiates (-), Cocaine (-), Amphetamines (-): Five reddish-purple lines—one Control line at the C position and one each at the THC, OPI, COC, and AMP positions.
- THC (-), Opiates (-), Cocaine (-), Amphetamines (+): Four reddish-purple lines—one Control line at the C position and one line each at the THC, OPI, and COC positions; no line at the AMP position.
 - THC (+), Opiates (-), Cocaine (+), Amphetamines (-): Three reddish-purple lines—one Control line at the C position, one line each at the OPI and AMP positions; no lines at the THC and COC positions.
 - THC (-), Opiates (+), Cocaine (-), Amphetamines (-): Four reddish-purple lines—one Control line at the C position and one line each at the THC, COC, and AMP positions; no line at the OPI position.
 - There are other possible results, depending on the combinations of drugs present in the urine sample.

Note: A very faint line for a specific drug in the result window, visible in 10 minutes, indicates that the amount of drug in the sample is near or below the cutoff level of the test. These urine specimens must be retested, or confirmed with a more specific alternative method such as gas chromatography/mass spectrometry, before positive determinations are made.

Limitations

- The test is designed for use with unadulterated urine only.
- There is a possibility that factors such as technical or procedural errors, as well as other substances in the urine sample than those listed in Table 4 below, may interfere with the test and cause erroneous results.
- Adulterants, such as bleach and/or alum, in urine specimens may produce erroneous results regardless of the method of analysis. If adulteration is suspected, the test should be repeated with a new sample.
- This test detects only the presence of THC metabolites, opiates, cocaine metabolites, amphetamines, and/or their metabolites in urine. A positive test result does not provide any indication of the level of intoxication or urinary concentration.
- The test result read after 10 minutes may not be consistent with the original reading obtained within the 10 minute reading period. The test result must be read within 10 minutes of sample application.
- Certain medications containing opiates or opiate derivatives, amphetamines, or methamphetamines may produce a positive result in any chemical or immunological assay. Additionally, foods and tea containing poppy products and/or coca leaves may produce a positive result. Prolonged passive smoking of THC may also produce a positive result.

User Quality Control

Quality Control. Control standards are not supplied with this kit; however, it is recommended that a control be tested as good laboratory testing practice. NIDA recommends that positive quality control specimens be at or near the cutoff concentration. For information on how to obtain controls, contact PBM's Technical Services. Before using a new kit with patient specimens, positive and negative controls should be tested to confirm the test procedure, and to verify the tests produce the expected Q.C. results. Q.C. specimens should also be run anytime there is any question concerning the validity of results obtained.

Process Control. The Control line can be considered an internal process control. A distinct reddish-purple Control line should always appear if the test procedure is performed properly, an adequate sample volume is used, the sample and reagent are wicking on the membrane, and the test reagents are working. If the Control line does not appear

in the control or validation line area, the test is invalid and a new test should be performed. If the problem persists, contact PBM for technical assistance.

Expected Values

AccuSign™ DOA 4 is a qualitative assay. The amount of drugs and metabolites present in urine cannot be estimated by the assay. The assay results distinguish positive from negative samples. Positive results indicate the samples contain the specific drug above the cutoff concentration.

Performance Characteristics

The **AccuSign™ DOA 4 Panel Assay** detects THC, opiates, cocaine, amphetamines, and their metabolites at cutoff levels based on the recommendations of the National Institute on Drug Abuse (NIDA) for screening of these drugs in urine.^{1,2,3}

THC	11-nor- Δ^9 -THC-9-carboxylic acid	50 ng/mL
OPI	Morphine	300 ng/mL
COC	Benzoyllecgonine	300 ng/mL
AMP	D-Amphetamine	1000 ng/mL

The accuracy of **AccuSign™ DOA 4** was evaluated in comparison to a commercially available immunoassay (Syva™ EMIT™ II) for each of the four drugs. About 1000 random clinical samples for each drug, including at least 250 positive samples above the cutoff level for each of the four drugs, was tested by both procedures, using the cutoff values listed. Complete agreement was observed in > 99% of the samples as shown below (Table 1.)

Table 1. Accuracy: Comparison of AccuSign™ DOA 4 with Syva™ EMIT™ II Assay

		Syva™ EMIT™ II (THC)		
		Positive	Negative	TOTAL
AccuSign™ DOA 4 (THC)	Positive	305*	5	310
	Negative	11	680	691
TOTAL		316	685	1001

		Syva™ EMIT™ II (OPI)		
		Positive	Negative	TOTAL
AccuSign™ DOA 4 (OPI)	Positive	249	0	249
	Negative	1	716	717
TOTAL		250	716	966

		Syva™ EMIT™ II (COC)		
		Positive	Negative	TOTAL
AccuSign™ DOA 4 (COC)	Positive	362	1	363
	Negative	14	644	658
TOTAL		376	645	1021

		Syva™ EMIT™ II (AMP/MET)		
		Positive	Negative	TOTAL
AccuSign™ DOA 4 (AMP)	Positive	185	0	185
	Negative	4	291	295
TOTAL		189	291	480

	Relative Sensitivity	Relative Specificity
THC	96.5% (305/316)	99.2% (680/685)
Opiates	99.6% (249/250)	> 99% (716/716)
Cocaine	96.3% (362/376)	99.8% (644/645)
Amphetamine	97.8% (185/189)	> 99% (291/291)

In a separate study, **AccuSign™ DOA 4** was evaluated against specimens confirmed as positive by GC/MS, for each of the four drugs. The results below demonstrate the excellent correlation of **AccuSign™ DOA 4** with GC/MS (99% agreement). (Table 2.)

Table 2. Accuracy: Comparison of AccuSign™ DOA 4 with GC/MS Assay

		AccuSign™	GC/MS
THC	Positive	87	88
	Negative	1	0
OPI	Positive	73	74
	Negative	1	0
COC	Positive	77	78
	Negative	1	0
AMP	Positive	55	56
	Negative	1	0

Precision and Accuracy

The precision of the AccuSign™ DOA 4 Panel Assay was determined by carrying out the test with serially diluted standard drug solutions. About 98% of the samples containing cocaine, opiates, or amphetamine and about 90% of the samples containing THC concentrations 25% over the cutoff level consistently showed positive results.

The study also included over 40 samples \pm 25% cutoff level as a challenge of cutoff precision. These results were found to be consistently in agreement with expected test results.

Distribution of Random Error:

Twenty (20) blind samples prepared by spiking various concentrations of cocaine, THC, morphine, or amphetamine were separately tested by two operators. The test results from the two operators showed complete agreement.

Reproducibility

The reproducibility of the test results of the AccuSign™ DOA 4 Panel Assay was examined at three different sites using a total of 15 blind controls, consisting of 5 negative samples, 5 moderately positive samples, and 5 strongly positive samples (i.e., a concentration 3 times the cutoff level). The results obtained at these three sites with these controls demonstrated 100% agreement with each other.

Specificity

The following table lists compounds that are detected by the AccuSign™ DOA 4 test. The specificity of the AccuSign™ DOA 4 test was determined by adding various drugs and drug metabolites to drug-negative urine specimens and testing with the AccuSign™ DOA 4 test kit. The results are expressed in terms of the concentration required to produce a positive result. (Table 3.)

Table 3. Specificity

Compound	Concentration (ng/mL)	% Cross-reactivity
THC		
Cannabinol	15,000	0.3
11-nor- Δ^9 -THC-9-COOH	30	100
11-nor- Δ^9 -THC-9-COOH	30	100
Δ^9 -THC	25,000	0.2
Δ^9 -THC	10,000	0.5
OPI		
Cocaine	300	100
Glucuronide	300	100

Hydrocodone	500	60
Hydromorphone	600	50
Levorphanol	5,000	6
Meprobamate	80,000	0.4
Morphine	300	100
Morphine-3- β -D-glucuronide	500	60
Nalorphine	1,000	30
Naloxane	100,000	0.3
Norcodone	60,000	0.5
Oxycodone	20,000	1.5
Oxymorphone	60,000	0.5
Procaine HCl	100,000	0.3
Theraine	5,000	6
COC		
Benzoylgonine	300	100
Cocaine HCl	500	60
Ergonine HCl	1,000	30
AMP		
D-Amphetamine	1,000	100
L-Amphetamine	7,000	14
D,L-Amphetamine sulfate	1,000	100
p-GH-Methamphetamine	30,000	3.3
Methylenedioxymethamphetamine	500	200
Methylenedioxyamphetamines	10,000	10
8-Phenethylamine	20,000	5
Phentermine	5,000	20
Trypamine	100,000	1
3-OH-Tyramine	90,000	1.1

The following compounds show no cross-reactivity when tested with AccuSign™ DOA 4 at a concentration of 100 μ g/mL. (Table 4.)

Table 4. Non Cross-Reacting Compounds

Acetaminophen	Dezoxipropoxyphene	Naproxen
Acetylsalicylate	Diazepam	Norethandrone
Amnopyrine	Diphenylhydantoin	Penicillin
Amiripyrine	Ephedrine	Pseudoephedrine
Amobarbital	Erythronium	Phenylhydrazine
Amoxapine	Etanol	Phenothiazine
Ampicillin	Estroic acid	Phenylpropanolamine
Apomorphine	Glutathione	Prednisone
Ascorbic acid	Guaiacal glycerol ether	Secobarbital
Atropine	Hydrocortisone	Tetracycline
Benzocaine	Isopropylalcohol	Tetrahydrozoline
Bisacarbital	Lidocaine	Trifluoperazine
Chlorazepoxide	Methadone	Tryptamine
Chlorpheniramine	Methaqualone	Zomepirac
Chlorpromazine	Methylphenyl	
Chloroquine		

References

- Hawkins RL, Chan CN, eds. *Urine Toxicology for Drugs of Abuse*. National Institute on Drug Abuse (NIDA) Research Monograph 73; 1986.
- Tierz, Norbert W. *Textbook of Clinical Chemistry*. W.B. Saunders Company 1986, p. 1735.
- Stewart DJ, Inaba I, Duncan M, and Kalow W. *Clin. Pharmacol Ther.* 1979;25: 21-4.
- Ambre JJ. *Anal. Toxicol.* 1983;9: 241-5.
- Basell RC. *Disposition of Toxic Drugs and Chemicals in Man*, 2nd Ed., Davis, CA: Biomedical Publ.; 1982. p.488.
- Blum K. *Handbook of Abusable Drugs*, 1st ed. New York: Gardner Press, Inc.; 1984.

PBM

Princeton BioMeditech Corporation
P.O. Box 7139, Princeton, New Jersey 08543-7139, U.S.A.
4242 U.S. Route 1, Monmouth Junction, New Jersey 08852-1905 U.S.A.
Tel: (908) 274-1000
Fax: (908) 274-1010

AccuSign™ is a Trademark of Princeton BioMeditech Corporation.

Patent Pending

© 1996 PBM
Printed in U.S.A.
P-5843-00 0218WK

AccuSign™ BAR

New One-Step Barbiturates Test

For In-Vitro Forensic Use Only

Simple One-Step Immunoassay for the
Qualitative Detection of Barbiturates in Urine

PBM

Catalog No. 7426-PBM 10 Test Kits
7426-PBM 10 10 Test Kits

Intended Use

The AccuSign™ BAR test is a simple, sensitive, immuno-
chromatographic assay for the rapid, qualitative detection of
barbiturates in urine with a result in 100 seconds.

The AccuSign™ BAR test provides only a preliminary
diagnostic result. A more specific, qualitative chemical
method must be used in order to obtain a confirmed
analytical result. Gas chromatography, mass spectrometry
(GC/MS) is the preferred confirmatory method. Other
chemical confirmatory methods are available. Clinical
confirmation and professional judgment should be ap-
plied to any drug of abuse test result, particularly when
preliminary positive results are seen.

Summary and Principle of Procedure

Barbiturates are a group of chemicals derived from barbi-
tic acid. Therefore, as barbiturates, they possess the central
nitrogen system. Taken orally in pill or tablet form, they are
prescribed for many medical conditions, usually for their
sedative effect. Abused barbiturates can, however, lead to
respiratory and renal system complications, and even death. The chemical
nature of barbiturates and alcohol is potentially dangerous.
Symptoms of barbiturate abuse include decreased alertness,
speech and memory. Acute abuse may include respiratory
collapse and loss of consciousness. Chronic use can re-
sult in severe alcoholism, seizures, and death. The effec-
tive half-life is 10-20 hours for phenobarbital. Barbi-
turates are usually present in detectable amounts for 1-2 weeks after
a 10-day use period (1).

Principle

The AccuSign™ BAR test uses solid phase immunoassay
technology for the qualitative detection of acetaldehyde and
barbiturate metabolites in human urine. The kit is based on the
principle of the highly specific enzyme reactions of reac-
tions between antigens and antibodies which have been used for the
analysis of specific substances in biological fluids. The test
relies on the competitive binding of the antibodies be-
tween competing conjugate antibodies to antibodies present in the
test sample. In this case, the conjugate antibody is labeled with
a color reagent. The color reagent is a sample of urine is
added to the sample. The color reagent is a color reagent to
increase the amount of color in the sample. The color reagent is
a color reagent to increase the amount of color in the sample.
The color reagent is a color reagent to increase the amount of
color in the sample. The color reagent is a color reagent to
increase the amount of color in the sample. The color reagent is
a color reagent to increase the amount of color in the sample.

Materials Provided

The AccuSign™ BAR test kit contains all the reagents
necessary to perform the assay.

- AccuSign™ BAR device. The test device consists of
a membrane in a well with a conjugate and a sample
well. The conjugate is a barbiturate antibody, the sample
is a sample of urine.
- Disposable sample diluent.
- Instructions for use.

Precautions

- Do not use the test kit for any other purpose.
- Avoid contact with the test kit. Avoid contact with the
test kit. Avoid contact with the test kit. Avoid contact with the
test kit.
- The test kit does not contain any hazardous materials.

EXHIBIT 2

Note: A very faint line on the Test window (TW) visible in 10 minutes indicates that the sensitivity of the assay in this sample is near or below the cutoff level of the test. These assay specimens must be treated as confirmed with a more specific alternative method such as gas chromatography/mass spectrometry before positive determinations are made.

Limitations

- The test is designed for use with adulterated urine only.
- There is a possibility that factors such as normal or physiological as well as other substances in the urine sample which are not listed in Tables 2 or 3 below may interfere with the test and cause erroneous results.
- Adulterants, such as bleach and/or other substances in some specimens may produce erroneous results regardless of the method of analysis. If adulteration is suspected, the test should be repeated with a new sample.
- This test detects only the presence of the biological adulterant determinants in urine. A positive test result does not provide any indication of the level of adulteration or urinary concentration.
- The test result (read after 10 minutes) may not be consistent with the original reading obtained within the 10 minute reading period. The test must be read within 10 minutes of sample application.
- Certain medications containing barbiturates may produce a positive result in any chemical or immunochemical assay.

User Quality Control

Quality Control: Control standards are not supplied with this kit. However, it is recommended that a control be tested in each laboratory using this protocol. NIDA recommends that a positive quality control specimen be used with the cutoff concentration. For information on how to obtain controls, contact PBMA Technical Services. Before using a new lot of reagent strips, strips, and/or control specimens, strips should be tested to confirm the lot procedure and to verify the test produces the expected Q/C results. Q/C specimens should also be run anytime there is any question concerning the ability of results obtained.

Positives Control: The Control line can be simulated as an internal process control. A distinct visible-purple Control line should always appear if the test procedure is performed properly, on adequate sample volume tested, the sample and reagent are working on the mechanism, and the test reagents are working. If the control line does not appear in subsequent results, the test area, the device or the test method should be performed. If the problem persists, contact PBMA for technical assistance.

Interference

Acetaminophen: Acetaminophen is a qualitative assay. The amount of acetaminophen or barbiturate metabolites present in the urine sample is estimated by the assay. The assay results do not distinguish between positive and negative samples. Positive results indicate the samples contain barbiturates above the cutoff concentration.

Performance Characteristics

The Acetaminophen BAR test has been shown to detect an average of 200 ng/mL of acetaminophen in urine. The test also detects only barbiturates found below the maximum concentration indicated (Table 2).

The accuracy of Acetaminophen BAR was evaluated on comparison specimens with available chemical assays (SP and PBM IT®). It is correct 100 samples was analyzed in a blind manner. The overall accuracy of the test was 98.3%, as shown below (Table 1).

Table 1. Acetaminophen/Barbiturate Acetaminophen BAR with SP and PBM IT

	SP vs. PBM IT (BAR)			
	Correct		Negative	
Acetaminophen BAR	Positive	105	0	105
	Negative	4	193	197
TOTAL		109	193	302

Acetaminophen BAR: 98.3% (108/110) = 98.3% (108/110)

Precision and Accuracy

The precision of Acetaminophen BAR was determined by testing ten test strips with each of three defined standard drug solutions. About 99% of the samples containing drug levels 25% or less of the cutoff level consistently showed positive results.

The study also included six or 40 samples at 25% cutoff level or in a change of cutoff precision. These results were found to be consistently in agreement with expected test results.

Distribution of Random Errors

Twenty CME blood samples, prepared by spiking various levels of amounts of drug were separately tested by two operators. The test results from the comparison showed complete agreement.

Reproducibility

The reproducibility of the test results of Acetaminophen BAR were assessed in terms of different screening a total of 14 blind samples, including 4 negative samples, 4 weak/trace positive samples, and 6 sample positive samples (e.g., 2 positive samples, 4 sample positive samples) at the cutoff level. The results showed that the test was consistent in determining 100% of the positive samples.



Presents the AccuSign™ DOA Series
One-Step Drug Test with
Results in Only 2-5 Minutes



- Easy to Read Color
- Highly Sensitive
- Built-In Test Control
- No Refrigeration
- Eliminates Timing
- Simultaneous Testing Capability of Multi-Drugs

Tests are available in single or multiple test panels for the following drugs:

Amphetamines

Cocaine

PCP

Barbiturates

Methamphetamines

THC (Marijuana)

Benzodiazepines


Morphine/Opiates

Watch us for More Tests & New Technology

EXHIBIT 3

AccuSign™ DOA Series


Test Procedure




1. Using the plastic pipette, add 3 drops of urine sample to the Sample Well.
2. Read results in 2 to 5 minutes (within 10 minutes).
3. Interpret Test Results Window:
 - **CONTROL LINE** - A colored line indicates the test is complete and the system has worked properly.
 - **NEGATIVE** - A colored line for the specific drug indicates the test is negative and the drug was NOT DETECTED.
 - **POSITIVE** - No colored line for the specific drug indicates the test is positive and the drug was DETECTED.

Tests Manufactured by Princeton BioMotech Corporation

Samples



THC (-) negative
 Opiates (-) negative
 Cocaine (-) negative
 Amphetamines (-) negative



THC (+) positive
 Opiates (-) negative
 Cocaine (+) positive
 Amphetamines (-) negative

For information or to place an order call:

UNIVERSAL DRUG TESTING
 476 ROUTE 61
 LARGE, PA 18026

Visualine™ II

One Step Drug Screening Test

AVITAR TECHNOLOGIES, INC.
produces, markets and distributes medical devices for
the health care industry.

Avitar now distributes rapid Tests for Drug Abuse
that are:

- accurate
- sensitive
- simple to use

The Visualine™ II tests¹ which have FDA approval to
market are:


- Cocaine
- Morphine
- Cannabinoids (THC)
- Benzodiazepines

The tests are composed of preformulated dry reagents
arranged on a porous membrane support in a convenient
cassette. To perform the tests, just add a few drops of urine to
the sample well of the device and wait five minutes. Results
are then easily read in the results window as the presence or
absence of a red line. A built in reference control ensures that
the sample has been added and that the test is effective. The
tests are based on the newest lateral flow micro-particle
immunoassay technology and have been evaluated in clinical
trials at a major university. The Visualine™ II tests are
designed to meet NIDA proposed cutoff levels.

¹ A test for Methamphetamine is available for uses that do not require FDA
approval to market.



Visualine™ II test device and pipette.



Simple to use and simple to read.

For information contact customer service at
1.800.255.0511

AVITAR TECHNOLOGIES, INC.
65 Dnn Road, Canton, Massachusetts 02021

EXHIBIT 4




EXHIBIT 5

Jun 21 1999 00:29:41 P

PHONE NO : 415 284 5098

Person : 1-STEP DETECT H55071947ES

For information about other home drug tests call 800-788-5716.
 Worldwide Medical Corporation, Irvine, CA 92618 • <http://www.wmcd.com>

EASY TO TEST



- Add 3 drops of urine
- Wait 5 minutes
- Read results

EASY TO READ RESULTS



First Check® provides immediate information about the use of marijuana, cocaine, opiates and amphetamines. The test strip, however, does not detect alcohol, benzodiazepines, barbiturates, butyrobacarbamate, or tricyclic antidepressants. For diagnosed and suspected users, the test strip can be used to monitor the effectiveness of treatment with these substances. This test contains:

- Test strip
- Sample dispenser
- Instructions

Manufactured by First Check®
 45281-06905



For further use only, enter in 45281-06905. Read enclosed directions, completely before use.

Commonly Asked Questions:

- A.** The test is not used to be the same shade of intensity. The color on the test card is a dark blue, or very weak.
- B.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- C.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- D.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- E.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.

Instructions for Use



First Check
 Marijuana (THC),
 Morphine/Opium,
 Cocaine, &
 Methamphetamine

When the need to know is... Now

- Simple - one step
- Easy-to read
- Confidential
- Result in 5 minutes

Read the following directions completely before use.

For external use only.

Store at 40-60°F (4-30°C).

FOR EDUCATIONAL USE ONLY

This Check is a registered trademark of First Check Medical, Inc.

Current Drug Usage Trends
 Marijuana use increased 5.5% in 1995, while heroin use fell 1.5% and cocaine use fell 1.5%. The National Institute on Drug Abuse (NIDA) reported that in 1995, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.


Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

Drug Abuse Statistics
 In 1994, cocaine-related deaths comprised 27% of all drug-related deaths. The National Institute on Drug Abuse (NIDA) reported that in 1994, 1.5 million people used drugs, down from 1.6 million in 1994.

WORLDWIDE MEDICAL
 1000 N. 10th St.
 Suite 100
 Phoenix, AZ 85016, U.S.A.
 Tel: 602.955.1000
 Fax: 602.955.1001
 E-mail: info@worldwide-med.com

- A.** The test is not used to be the same shade of intensity. The color on the test card is a dark blue, or very weak.
- B.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- C.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- D.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- E.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.

- A.** The test is not used to be the same shade of intensity. The color on the test card is a dark blue, or very weak.
- B.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- C.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- D.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.
- E.** The test should be held within 10 minutes for best results. A negative result is best, 1 line in the test window and 4 lines in the control window are best.



JUN 21 1999 09:35AM PS

PHONE NO. : 412 384 5260

EP004 : 1-STEP DETECT ASSOCIATES

For information about other home drug tests call 866-788-5716.
 Worldwide Medical Corporation, Irvine, CA 92618 • <http://www.wmed.com>

EASY TO TEST



- Add 3 drops of urine
- Wait 5 minutes
- Read results

EASY TO READ RESULTS



First Cause provides immediate information about the use of substances and enables it to act for legal, law enforcement, or medical purposes. For diagnosis and treatment, consult with health care or laboratory testing professionals.

This box contains:
 1 test unit
 1 sample dispenser
 1 sample collection tube
 1 sample collection tube
 1 sample collection tube



For results see web site at www.wmed.com or call 866-788-5716. Read results immediately, do not use.

First Check[®] Marijuana & Cocaine


May be the *only* microplate



Marijuana & Cocaine

THC is the primary active ingredient in marijuana (cannabis). When inhaled or smoked, it produces euphoric effects. Users have impairment of short term memory and attention are often learning. Also, it may cause transient episodes of confusion, anxiety, or even frank toxic delirium. Long term, relatively heavy use may be associated with behavioral disorders. The peak effect of smoking marijuana occurs in 20-30 minutes and the duration is 95-120 minutes after one cigarette. The stable levels of urinary metabolites are found within hours of exposure and remain detectable for 3-10 days after smoking.

Cocaine derived from the leaves of coca plant is a potent central nervous system (CNS) stimulant and a local anesthetic. Cocaine induces euphoria, confidence, and a sense of increased energy in the user; these psychological effects are accompanied by increased heart rate, dilation of the pupils, fever, tremor and sweating. Cocaine is used by smoking, intravenous, intranasal or oral administration and excreted in the urine primarily as benzoylecgonine in a short time. Benzoylecgonine



is a weak, non-opioid analgesic (2-3 times as strong as cocaine) 5-15 hours and can specifically be detected for 20-40 hours after cocaine use or exposure.

Before you begin

Read all the information in this pamphlet before performing the test. First, make sure you are familiar with the test kit contents shown below. Store at 35-56° F (2-3° C) in the sealed pouch, away from direct sunlight. Do not use after the expiration date stamped on the package.

Instructions

1. Open the sealed pouch, remove the First Check[®] card, and set the card on a flat surface with Test and Control windows facing up.

2. Collect urine sample in a clean plastic or glass container.

3. Withdraw dispenser set over sample. Press bulb between thumb and index finger, then dispenser opening into sample and release pressure on bulb. Sample will fill half of dispenser tube.



4. With sample dispenser in vertical position over Sample well of test card, gently squeeze dispenser bulb to allow 3 fully-formed drops of urine, one at a time, to fall into Sample well.



5. Allow the test card to remain unattended until result is read. Read the result after 5 minutes but within 10 minutes.



Results

2-5 or 3+ lines: Test and result show 10 minutes after testing result.

Negative

Three horizontal lines: one line in the Control (upper) window, and two lines in the Test (lower) window, means there is no marijuana and no cocaine present in the urine sample. The line at the Test is an indicator that the liquid or absorbent line in the Control window.



Positive

One line in the Control window and no line in the Test window: no line in the Test window means the sample contains no drug.



Invalid Test

A distorted colored line should always appear in the Control (upper) window. If no line appears in the Control window, do not interpret result.



Limitations

The First Check[®] Line-Stop Home Drug Test is not available. The user must be of legal age. The test detects only the presence of marijuana (THC) and cocaine or their metabolites in urine. A positive test does not provide any information about the amount or level of intoxication.

The test is designed for use with undiluted urine, only. Adulterants, such as bleach or salt, may produce a false negative result. If adulteration is suspected, the test should be repeated with a new urine sample.

The result may be read 5-10 minutes after sample application. A result read after 10 minutes may not be accurate. Urine sample should be at room temperature. If sample has been refrigerated, allow sample to come to room temperature before testing.

Physiological exposure to secondhand marijuana smoke may produce a positive result.



Jun 21 1999 09:43:41 P11

PHONE NO. : 412 384 5560

FROM : 1-STEP DETECT ASSOCIATES

For information about other home drug tests call 800-288-5716.
 Workwide Medical Corporation, Irvine, CA 92618 • <http://www.wmmed.com>

EASY TO TEST



- Add 3 drops of urine
- Wait 5 minutes
- Read results

EASY TO READ RESULTS



Just Check! No flame, no toxic fumes, no odor, no waiting. Results in 5 minutes. For diagnosis and treatment, consult with health care or substance abuse professional.

This box contains:

- 1 test unit
- 1 sample container
- 1 sample exposure
- 1 instruction

Material required but not provided:

- One glass or plastic container for collection of sample



For a detailed use and safety guide, visit www.wmmed.com. Please read and follow directions completely before use.

First Check[®] Marijuana (THC)

Not to be taken internally.




Marijuana (THC)

THC is the primary active ingredient in marijuana (cannabis). When inhaled or smoked, it produces euphoric effects. Users have impairments of short-term memory and marijuana use slows learning. Also, it may cause transient episodes of confusion, anxiety, or even faint, toxic delirium. Long-term, relatively heavy use may be associated with behavioral disorders. The peak effect of smoking marijuana occurs in 20-30 minutes and the duration is 90-120 minutes after one cigarette. Elevated levels of urinary metabolites are found within hours of exposure and remain detectable for 3-10 days after smoking.

Before you begin

Read all the information in this pamphlet before performing the test. First, make sure you are familiar with the test kit contents shown, below. Score at 36-65°F.



2-10°C in the sealed pouch, away from direct sunlight, to not see after the expiration date, stamped on the package.

Instructions

1. Open the sealed pouch, remove the First Check[®] card, and set the card on a flat surface with Test and Control windows facing up.
2. Collect urine sample in a clean plastic or glass container.



3. With sample dispenser over sample, press bulb between thumb and index finger, insert dispenser opening into sample, and release pressure on bulb. You should see sample fill half of dispenser tube.

Do not discard the unused urine, until after the test has been completed and the result interpreted.



4. With sample dispenser in vertical position over Sample well of test card, gently squeeze dispenser bulb to allow fully formed drop of urine, one at a time, to fall into Sample well.



5. Allow the test card to remain undisturbed until result is read. Read the result after 3 minutes but within 10 minutes.



Results

At least 2 minutes but not more than 10 minutes before reading result.



2 lines - negative

Two horizontal lines, one in each of the Control (upper) window and Test (lower) window, means there is no marijuana present in the urine sample. The line in the Test window may be lighter or darker than the line in the Control window.

1 line - positive

One line in the Control window and no line in the Test window means the sample contains marijuana.



Invalid Test

A distinct colored line should always appear in the Control (upper) window. If no line appears in the Control window, do not interpret result.

Limitations

The First Check[®] One-Step Home Drug Test is used as a screening tool. The test instructions must be followed precisely.

The test detects only the presence of marijuana (THC) in metabolites in urine. A positive test (no line in Test window) does not provide any information about the amount or level of intoxication.

The test is designed for use with unadulterated urine. Only adulterants, such as diluted water, alter a urine sample may produce an erroneous result. If adulteration is suspected, the test should be repeated with a new urine sample. The result must be read 3-10 minutes after sample application. A result read after 10 minutes may not be accurate.

Urine sample should be at room temperature. If sample has been refrigerated, allow sample to come to room temperature before testing.

Prolonged exposure to secondhand marijuana smoke may produce a positive result.